

▶ Acurex™ compact 501



- 1 | Integrated dosing mechanism. Yesmple and robust construction.
- 2 | Retractable graduated column reduces the height of the instrument
- 3 | The fluid is not in contact with metals along its flow
- 4 | Four container sizes.
- 5 | Fully autoclavable at 121°C / 250°F.

Code	Volume	Divisions	Accuracy (%CV)			Capacity*
			Vol. min	Vol. medium	Vol. max	
MDD005	1 – 10 mL	0.2 mL	<0.5%	<0.35%	<0.1%	1000 mL
MDD006	1 – 10 mL	0.2 mL	<0.5%	<0.35%	<0.1%	2000 mL
MDD007	0.2 – 2 mL	0.1 mL	<0.5%	<0.35%	<0.1%	1000 mL
MDD008	0.2 – 2 mL	0.1 mL	<0.5%	<0.35%	<0.1%	2000 mL
MDD009	0.2 – 2 mL	0.1 mL	<0.5%	<0.35%	<0.1%	250 mL
MDD011	0.4 – 5 mL	0.2 mL	<0.5%1)	<0.35%	<0.1%	1000 mL
MDD012	0.4 – 5 mL	0.2 mL	<0.5%1)	<0.35%	<0.1%	2000 mL
MDD013	0.4 – 5 mL	0.2 mL	<0.5%1)	<0.35%	<0.1%	500 mL
MDD010	1 – 30 mL	1.0 mL	<0.5%2)	<0.35%	<0.1%	2000 mL

\* of the container

## Dispenser Calibrex™ 520

- 1 | Increased chemical resistance. Stable performance over time.
- 2 | Yesmple construction - no tools required for disassembly.
- 3 | Immediate volume adjustment.
- 4 | Laboratory calibration.
- 5 | Autoclavable at 121°C / 250°C°CF fully assembled.
- 6 | Three bottleneck adapters included.



*It presents an easy volume adjustment*



Code	Volume	Divisions	Accuracy (%E)			Accuracy (%CV)		
			Vol. min	Vol. medium	Vol. max	Vol. min	Vol. medium	Vol. max
MDD014	0.25-2 mL	0.05 mL	<±3.0%	<±1.8%	<±0.6%	<0.5 %	<0.35 %	<0.1%
MDD015	1-5 mL	0.1 mL	<±2.0%	<±1.3%	<±0.6%	<0.5 %	<0.35 %	<0.1%
MDD016	1-10 mL	0.25 mL	<±1.5%	<±1.1%	<±0.6%	<0.5 %	<0.35 %	<0.1%

Organo 525

Code	Description	Volume	Divisions	Accuracy (E%)			Accuracy (E%)		
				Vol.min	Vol medium	Vol max	Vol. min	Vol medium	Vol max
MDD017	Estándar	0.1-1 mL	0.02 mL	<± 3,0%	<±1,8%	<±0,6%	<1,2%	<0,7%	<0,17%
MDD018	Estándar	0.25-2.5 mL	0.05 mL	<±2,7%	<±1,6%	<±0,6%	<0,9%	<0,55%	<0,17%
MDD019	Estándar	0.5-5 mL	0.1 mL	<±2,0%	<±1,3%	<±0,6%	<0,5%	<0,35%	<0,1%
MDD020	Estándar	1-10 mL	0.2 mL	<±1,5%	<±1,2%	<±0,6%	<0,5%	<0,85%	<0,1%
MDD024	Estándar	2.5-25 mL	0.5 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD026	Estándar	5-50 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD028	Estándar	10-100 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD021	Con llave	0.1-1 mL	0.02 mL	<± 3,0%	<±1,8%	<±0,6%	<1,2%	<0,7%	<0,17%
MDD022	Con llave	0.25-2.5 mL	0.05 mL	<±2,7%	<±1,6%	<±0,6%	<0,9%	<0,55%	<0,17%
MDD023	Con llave	0.5-5 mL	0.1 mL	<±2,0%	<±1,3%	<±0,6%	<0,5%	<0,35%	<0,1%
MDD030	Con llave	1-10 mL	0.2 mL	<±1,5%	<±1,2%	<±0,6%	<0,5%	<0,85%	<0,1%
MDD025	Con llave	2.5-25 mL	0.5 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD027	Con llave	5-50 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD029	Con llave	10-100 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%

Solutae 530

Code	Description	Volume	Divisions	Accuracy (E%)			Accuracy (E%)		
				Vol.min	Vol medium	Vol max	Vol. min	Vol medium	Vol max
MDD031	Estándar	0.1-1 mL	0.02 mL	<± 3,0%	<±1,8%	<±0,6%	<1,2%	<0,7%	<0,17%
MDD032	Estándar	0.25-2.5 mL	0.05 mL	<±2,7%	<±1,6%	<±0,6%	<0,9%	<0,55%	<0,17%
MDD033	Estándar	0.5-5 mL	0.1 mL	<±2,0%	<±1,3%	<±0,6%	<0,5%	<0,35%	<0,1%
MDD034	Estándar	1-10 mL	0.2 mL	<±1,5%	<±1,2%	<±0,6%	<0,5%	<0,85%	<0,1%
MDD038	Estándar	2.5-25 mL	0.5 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD040	Estándar	5-50 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD042	Estándar	10-100 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD035	Con llave	0.1-1 mL	0.02 mL	<± 3,0%	<±1,8%	<±0,6%	<1,2%	<0,7%	<0,17%
MDD036	Con llave	0.25-2.5 mL	0.05 mL	<±2,7%	<±1,6%	<±0,6%	<0,9%	<0,55%	<0,17%
MDD037	Con llave	0.5-5 mL	0.1 mL	<±2,0%	<±1,3%	<±0,6%	<0,5%	<0,35%	<0,1%
MDD044	Con llave	1-10 mL	0.2 mL	<±1,5%	<±1,2%	<±0,6%	<0,5%	<0,85%	<0,1%
MDD039	Con llave	2.5-25 mL	0.5 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD041	Con llave	5-50 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%
MDD043	Con llave	10-100 mL	1.0 mL	<±1,5%	<±1,1%	<±0,6%	<±0,5%	<±0,35%	<±0,1%